

Polarity-specific changes in the E-field and focality in mild cognitive impairment patients for HD-tDCS and reverse HD-tDCS

Utkarsh Pancholi^{1,2}, Vijay Dave^{3,2,*}

ABSTRACT

Objective: This study analyzed polarity-specific changes in the electric field (E-field) and the focal point of tDCS stimulation for both High-Definition tDCS (HD-tDCS) and reverse HD-tDCS considering the head geometry of Mild Cognitive Impairment (MCI) patients. The distance between the left preauricular (LPA) and right preauricular points as well as theinion and nasion were calculated in 3D to measure correlations. **Methods:** T1-weighted structural MRI images were obtained from the Alzheimer's Disease Neuroimaging Initiative (ADNI). We utilized SimNIBS 3.2.6 for MRI image processing, 3D head model generation, application of HD-tDCS and reverse HD-tDCS, E-field analysis, and other relevant parameters. **Results:** Electric field analysis was conducted on all 10 MCI subjects to examine the distribution and focality of the peak electric field. Despite using uniform stimulation for all subjects, variation was observed in the peak electric field (at the 99th, 99th, and 95th percentiles) and focality (at the 75th and 50th percentiles). However, the peak electric field and focality values remained consistent when comparing HD-tDCS and reverse HD-tDCS. The only differing quantities were associated with the normal component influenced by the changing direction of the current in HD-tDCS and reverse HD-tDCS. **Conclusion:** HD-tDCS and reverse HD-tDCS exhibited distinct values for the electric field strength and focality. HD-tDCS demonstrated more positive fields and inflow current compared to reverse HD-tDCS, making it suitable for excitatory effects. Conversely, reverse HD-tDCS would be more appropriate for inhibitory effects due to the prevalence of more negative fields and outflow current.

Key words: HD-tDCS, Electric field strength, Focality, Mild cognitive impairment

¹Biomedical Engineering Department, L.D. College of Engineering, Ahmedabad, Gujarat, India

²Gujarat Technological University, Ahmedabad, Gujarat, India

³Research Supervisor, Biomedical Engineering Department, Government Engineering College, Sector-28, Gandhinagar, Gujarat, India

Correspondence

Vijay Dave, Research Supervisor, Biomedical Engineering Department, Government Engineering College, Sector-28, Gandhinagar, Gujarat, India

Gujarat Technological University, Ahmedabad, Gujarat, India

Email: vijaydave12@gmail.com

History

- Received: Aug 10, 2023
- Accepted: Feb 19, 2024
- Published Online: Feb 29, 2024

DOI : 10.15419/bmrat.v11i12.867



Copyright

© Biomedpress. This is an open-access article distributed under the terms of the Creative Commons Attribution 4.0 International license.



INTRODUCTION

Mild cognitive impairment (MCI) refers to a state of both physiological and psychological decline that affects cognitive and memory-related functions in the elderly population. According to the World Health Organization, an estimated 426 million individuals will be over the age of 60 by the year 2050¹. MCI, as a symptomatic neurodegenerative disorder, is categorized into amnesic (affecting short-term memory in the prefrontal cortex) and non-amnesic (impaired problem-solving, visual perception, language formation, attention, and concentration) conditions that progressively worsen over time. To diagnose and confirm the clinical presence of MCI, laboratory tests are conducted to measure various hormone levels (e.g., thyroid hormone²), examine blood and urine for infections³, and analyze cerebrospinal fluid⁴ to assess autoimmune conditions related to neurodegenerative diseases. Additional diagnostic approaches encompass imaging methods such as Magnetic Resonance Imaging (MRI)⁵, amyloid scans using Positron Emission Tomography (PET)⁶, and Computed Tomography (CT)⁷. Neuropsychological assessments such

as the Montreal Cognitive Assessment (MoCA) are also utilized for diagnosis⁸. Despite the absence of compelling evidence and approved medications for MCI treatment⁹, clinicians have proposed alternative therapies such as physical activity, dietary and nutritional interventions, and behavioral therapy¹⁰. Depression, anxiety, disinterest, and aggression are recognized as co-occurring symptoms in conjunction with the clinically observed manifestations of MCI¹¹. The symptoms of depression vary across age groups, with manifestations including sadness and irritability in young children, while memory issues, fatigue, and suicidal thoughts are more prevalent in adults and older individuals. When cognitive impairments (non-amnesic) accompany depression, the severity of the disease increases, and the likelihood of progression to dementia becomes more significant within five years of MCI onset¹².

The electrostimulation method known as transcranial direct current stimulation (tDCS) has gained popularity among psychiatrists and other medical practitioners specializing in the treatment of neurodegenerative conditions such as depression, MCI,

Cite this article : Pancholi U, V.D. Polarity-specific changes in the E-field and focality in mild cognitive impairment patients for HD-tDCS and reverse HD-tDCS. *Biomed. Res. Ther.* 2024; 11(2):6209-6223.

and Alzheimer's disease (AD). Early research findings have shown encouraging outcomes, particularly in the treatment of depression using anodal stimulation and addressing hyperactive brain disorders (e.g., mania) with cathodal stimulation^{13,14}. As a non-invasive therapeutic approach, tDCS modifies neuronal activity to produce its therapeutic effects, increasing the excitability of neuronal cells without inducing action potentials. The objective of this therapy is to modify brain function and improve cognitive abilities in patients with neurological and psychological conditions. tDCS utilizes a direct current of 1-2 mA delivered to the scalp through bipolar or multiple stimulating metal electrodes. The therapist has the flexibility to adjust the current intensity, location, duration, and polarity of electrodes based on the individual needs of the patient. Bipolar stimulation involves a two-electrode method where one electrode serves as the active stimulator (anode) placed on the target area of the brain, while the other electrode functions as the cathode and is typically positioned in an extracephalic region¹⁵. The placement of these electrodes determines the inward current (at the anode) and outward current (at the cathode), leading to depolarization and repolarization of the neural mass, respectively. The resistance of head tissues such as the skin, scalp, skull, white matter, and grey matter reduces the intensity of the current and the focusing of the electrical field as it penetrates deeper regions of the brain. This direct current, reaching various tissues of the head, affects the electric field strength (measured in V/m) in both the targeted area and its surrounding regions¹⁶. The electric field strength and focality are influenced differently by stimulation parameters such as the intensity¹⁷, size, and shape of the electrode¹⁸, the target location, the duration of stimulation, and the conductivity of head tissues¹⁹.

There are two methods of delivering currents to the head: bipolar stimulation and HD-tDCS. Bipolar stimulation involves the use of two stimulating electrodes (the anode and cathode), while HD-tDCS employs multiple electrodes (such as 3×1, three anodes and one cathode, or 4×1, four anodes and one cathode). When using bipolar stimulation on opposite hemispheres, with one electrode on the left and the other on the right hemisphere, the electric field strength becomes less focused due to the current flowing from the anode to the cathode and passing through multiple brain tissues. However, focality can be improved by using a finer diameter electrode located on the same hemisphere. In their study, Datta et al. conducted a comparison between conventional bipolar and HD-tDCS electrode configurations, and

the results revealed that the 4×1 HD-tDCS demonstrated increased focality and enhanced electric field strength in the targeted area compared to its counterpart²⁰. Gaurav Sharma *et al.* conducted simulations of bipolar stimulation using electrodes of varying sizes (4×4, 3×3, and 2×2 cm², with a sponge electrode thickness of 2.0 mm and a conductive gel thickness of 0.50 mm). They then compared these results with the HD-tDCS configuration (specifically, 4×1 with one anode and four cathodes) using four different electrode locations with Montreal Neurologic Institute coordinates (MNI)²¹. Their findings indicated that the focality was greater using the HD-tDCS electrode placement method compared to the bipolar configuration. Edgard Morya et al. conducted a review of the effects of transcutaneous spinal direct current stimulation (ts-tDCS) and cerebellar tDCS, and the results showed improvements in alleviating symptoms of neurological disorders, providing pain relief, and enhancing cognitive and rehabilitative motor functions when using HD-tDCS²². The majority of the research conducted on HD-tDCS has emphasized its focality compared to conventional bipolar tDCS, where all four electrodes are equidistant from the central electrode. To date, only anodal HD-tDCS stimulation has been reported, with findings related to the electric field strength and its spatial characteristics. However, there is insufficient data available regarding the effects of reversing the direction of the current (anodal to cathodal) for MCI patients.

In this research, we conducted simulations of both anodal and cathodal 4×1 HD-tDCS in MCI patients, targeting the frontal cortex to address cognitive decline with symptoms of depression. The study's objectives were as follows: 1) Creating a high-precision 3D head model using SimNIBS 3.2.6; 2) Simulating the 3D head model with the 4×1 anodal and cathodal HD-tDCS configurations; 3) Determining the patient-specific relative distance from the central electrode (anodal or cathodal); 4) Evaluating the relative distance between the Iz and Nz (inion and nasion) and the RPA and LPA (Right pre-auricle point and left pre-auricle point) to characterize the geometrical changes in patients; 5) Analyzing the incoming and outgoing current at the normal axis to the current stimulation.

METHODS

Computational methods are employed to visualize various characteristics of the electric field, such as the E-field strength, spatial distribution, intensity, current density, and their components at the normal axis. *In-situ* platforms such as COMSOL Multiphysics

and Solidworks are utilized for customizing 3D head modeling and E-field analysis. Additionally, dedicated software packages such as Simulation of Non-invasive Brain Stimulation (SimNIBS), Computation of the Electric field due to Transcranial current Stimulation (COMET), SPHEARE, and ROAST (a fully automated, realistic, volumetric approach to simulate transcranial electric stimulation) are available for tDCS modeling and analysis²³. For our study, we utilized SimNIBS 3.2.6 for MRI image processing, 3D head model generation, the application of HD-tDCS and reversed HD-tDCS, E-field analysis, and other relevant parameters. **Figure 1** illustrates the processing pipeline for the tDCS simulation employed in our work.

Structural MRI

Structural MRI enables the comprehensive visualization of the entire head anatomy, encompassing the dimensions, configuration, and integrity of different brain structures, including the cerebral cortex, white matter, grey matter, and deeper regions of the brain. MRI images are typically essential for detecting any structural abnormalities within the brain and are commonly used for prognosis and assessing the effects of treatment over time. Structural MRI scans employ T1 and T2-weighted imaging sequences. T1-weighted images are valuable for revealing anatomical structures with clear contrast, allowing differentiation between grey matter and white matter in the brain, identifying lesions or tumors, and evaluating normal tissue morphology^{24,25}. T2-weighted images are useful for examining conditions that entail fluid accumulation or edema, detecting issues such as inflammation, infection, or tumors, and assessing specific abnormalities such as white matter lesions in cases of multiple sclerosis²⁶.

We obtained T1-weighted structural images from the Alzheimer's Disease Neuroimaging Initiative (ADNI) for ten MCI patients (eight males and two females, median age: 71 years). The imaging protocol for all the images included the following specifications: acquisition plane=sagittal; acquisition type = coil =sense-head; field strength = 3.0 tesla; flip angle = 8.0 degrees; manufacturer = Philips Medical Systems; matrix x = 256.0 pixels; matrix y = 256.0 pixels; matrix z = 170.0; mfg. model = Intera; pixel spacing x = 1.0 mm; pixel spacing y = 1.0 mm; pulse sequence = GR; slice thickness = 1.2 mm; TE=3.2 ms; TI = 0.0 ms; TR=6.8 ms, and weighting = t1²⁷.

3D Head model

The process of 3D head modeling, also referred to as volume conductor modeling, was accomplished using the headreco segmentation pipeline within SimNIBS 3.2.6. This command line pipeline generates a finite element mesh (FEM) by allocating voxels to distinct head tissue classes. Headreco utilizes statistical parametric mapping (SPM 12) and computational anatomy toolbox (CAT 12) for the surface reconstruction of the brain's grey matter²⁸. The headreco generates 0.5 nodes per mm³ of volume, and this resolution can be adjusted as needed, either increased or decreased. For this study, ten head models were created corresponding to the ten MCI patients selected for analysis.

HD-tDCS placement

tDCS is a well-known therapeutic approach involving a conventional protocol with two electrodes (bipolar) stimulating the targeted brain area. These electrodes consist of one anode (positive) and one cathode (negative) with precise sizes and shapes. In our study, we employed HD-tDCS and reversed HD-tDCS configurations to investigate the changes in the electric field distribution and input/output current in the targeted brain region. Since depression is one of the symptoms of MCI patients, we selected the frontal lobe as the target area. Two stimulation protocols were used: 1) 4×1 anodal stimulation with one anode and four cathodes, and 2) 4×1 cathodal stimulation with one cathode and four anodes. We adopted the 10-20 electrode placement method used in electroencephalography to determine the positions of the stimulating electrodes. In the first protocol, F3 served as the anode, while AF3, F1, FC3, and F5 acted as cathodes. In the second protocol, F3 acted as the cathode, while AF3, F1, FC3, and F5 served as anodes. The electrode morphology was round, with a diameter of 1.0 cm and a thickness of 5 mm. The anode current was set at 2 mA, and the cathode current was set at 0.500 mA. We calculated the distance of each electrode from the central anodal or cathodal electrode. Additionally, the distance between theinion and nasion was measured using fiducial points. **Figure 2** illustrates the HD-tDCS placement on the head scalp (a) and the gray matter (b). **Table 1** and **Table 2** list the distances between theinion-nasion and LPA-RPA and the distances of each electrode from the central electrode, respectively.

Table 1: Distance between inion-nasion and LPA-RPA

Sr. No.	Subject ID	Sex	Age	Fiducials Points			Distance between Nz & Iz (mm)			LPA			Distance between LPA & RPA (mm)				
				X	Y	Z	X	Y	Z	X	Y	Z	X	Y	Z		
1	Sub_01	F	70	3.23	99.57	-14.85	185.98	-9.81	-84.43	-38.62	-74.39	10.48	-31.80	142.09	66.95	2.27	-43.86
2	Sub_02	M	83	-3.53	108.78	4.92	227.33	-14.72	-98.15	-88.51	-87.53	11.98	-50.79	159.97	72.21	3.48	-49.44
3	Sub_03	M	79	-7.22	116.35	-37.60	203.67	8.67	-59.17	-139.69	-74.49	21.73	-88.35	152.77	76.18	44.31	-99.63
4	Sub_04	M	69	-8.50	99.07	-11.61	187.06	-1.40	-87.74	-18.14	-77.67	1.66	-24.90	146.55	68.87	2.15	-25.64
5	Sub_05	F	55	2.64	84.50	-16.30	173.82	-4.85	-89.03	-9.54	-67.49	1.54	-22.67	134.74	66.75	-9.98	-21.72
6	Sub_06	M	75	2.13	110.66	6.39	204.90	-1.34	-82.67	-61.39	-79.93	14.11	-34.63	159.79	79.74	11.68	-40.48
7	Sub_07	M	56	7.60	105.32	-3.10	200.00	-3.45	-94.14	-12.81	-75.13	8.43	-15.18	154.98	79.44	-1.42	-20.70
8	Sub_08	M	67	3.72	110.92	-0.98	207.21	0.83	-90.39	-49.98	-78.34	9.01	-30.65	160.00	81.39	9.54	-39.85
9	Sub_09	M	77	1.23	130.20	-17.95	202.71	-1.85	-65.32	-71.40	-75.90	35.95	-59.09	154.44	78.24	28.95	-52.50
10	Sub_10	M	79	0.94	114.11	4.38	216.77	0.90	-96.90	-45.27	-78.47	9.12	-24.50	157.25	78.41	11.09	-35.11
Total distance							2009.45			1522.57							
Average distance							200.95			152.26							




Figure 1: Processing pipeline for tDCS simulation.




Figure 2: HD-tDCS electrode placement (F3-Anode, AF3/F1/FC3/F5- cathode) on scalp (a) and Gray matter (b).

Table 2: Distance between central electrodes to other four electrode

Sr. No.	Subject ID	Sex	Age	Distance between central electrode and other four electrodes (mm)			
				F3-AF3	F3-F1	F3-FC3	F3-F5
1	Sub_01	F	70	29.843503	24.238601	31.237139	24.156866
2	Sub_02	M	83	33.49835	27.007443	34.903726	26.785181
3	Sub_03	M	79	31.078774	25.95485	33.746791	23.92078
4	Sub_04	M	69	30.603582	26.031338	32.085855	23.97301
5	Sub_05	F	55	28.148209	24.3985	30.402423	23.962548
6	Sub_06	M	75	31.568736	27.063116	34.785291	26.114197
7	Sub_07	M	56	33.1169	26.292112	34.797158	26.098594
8	Sub_08	M	67	33.145992	28.121298	35.029023	26.144385
9	Sub_09	M	77	32.185293	27.183191	34.010225	26.331693
10	Sub_10	M	79	33.877842	27.504808	36.312091	25.536077
Total				317.067181	263.795257	306.072583	253.023331
Average				31.7067181	26.3795257	30.6072583	25.3023331

E-field analysis

To anticipate the stimulation effects and investigate potential side effects, the spatial extent and intensity of the electric field (E-field) in various brain regions were analyzed and visualized using gmsh. The governing equations for the electrical current flow in tDCS are rooted in the principles of electrostatics. Two fundamental equations are utilized: Ohm's law and Laplace's equation. Ohm's law links the electric current (I) flowing through a conductor to the voltage (V) across it and the resistance (R) it offers. In the context of tDCS, the current flows through the head and brain tissues, and Ohm's law can be locally applied at each point. The current (I) is expressed as $\sigma \cdot E$, where I represents the electric current density vector (A/m^2); σ is the conductivity of the medium (S/m), and E denotes the electric field vector (V/m). Laplace's equation characterizes the distribution of electric potential (V) in a region without charge density. In tDCS, the equation governs the behavior of the electric potential throughout the head and brain tissues. The electric potential ($\nabla^2 V$) is equal to zero, where ∇^2 represents the Laplacian operator ($\partial^2/\partial x^2 + \partial^2/\partial y^2 + \partial^2/\partial z^2$), and V signifies the electric potential (V). These equations, along with appropriate boundary conditions, serve as the foundation for the computational modeling of tDCS. The electric field (E-field) distribution can be derived from the electric potential (V) using the relationship $E = -\nabla V$, where ∇ signifies the gradient operator ($\partial/\partial x, \partial/\partial y, \partial/\partial z$)^{29,30}.

E-normal and J-normal component analysis


In tDCS, researchers frequently analyze the E-normal (electric field normal) and J-normal (current density normal) components to assess the impact of stimulation on neuronal activity and the potential for neural modulation. These components offer valuable information concerning the direction and strength of the electric field and current density in relation to the brain's surface. The E-normal component specifically refers to the electric field vector that is perpendicular to the brain's surface, indicating the direction and intensity of the electric field relative to the underlying neural tissue. Through the examination of the E-normal component, both the spatial extent and strength of the electric field at the brain's surface can be determined. This information is significant, as it aids in comprehending the areas most impacted by tDCS, enabling precise electrode placement to target specific brain regions. The J-normal component pertains to the current density vector perpendicular to

the brain's surface, indicating the flow of electrical current through neural tissue in that direction. Examining the J-normal component offers valuable information about the distribution and strength of the current density at the brain's surface, shedding light on how the electrical current is distributed within the brain tissue and its effects on neuronal activity.

RESULTS

Electric field analysis was conducted on all 10 MCI subjects to examine the distribution and focality of the peak electric field. Despite using uniform stimulation for all subjects, variation was observed in the peak electric field (at the 99.9th, 99th, and 95th percentiles) and focality (at the 75th and 50th percentiles). However, when comparing HD-tDCS and reversed HD-tDCS, the peak electric field and focality values remained constant (Table 3). The only differing quantities were associated with the normal component, which was influenced by the changing direction of the current in HD-tDCS and reversed HD-tDCS. The average peak electric field for the normal component, Norm E, was 1.62E-01 at the 99.9th percentile, 9.65E-02 at the 99th percentile, and 3.44E-02 at the 95th percentile. For the normal component, Norm J, the average peak electric field was 4.46E-02 at the 99.9th percentile, 2.65E-02 at the 99th percentile, and 9.48E-03 at the 95th percentile. The focality for Norm E was 3.00E+03 at the 75th percentile and 9.55E+03 at the 50th percentile; for Norm J, this was 3.00E+03 at the 75th percentile and 9.55E+03 at the 50th percentile.

Columns 6 to 15 in Table 4 and Table 5 display the normal components of the E vector representing the peak electric fields for E-normal (measured in V/m) and J-normal (measured in A/m^2). The letter "P" signifies positive values, indicating the incoming electric field entering the surface, while "N" denotes negative values, indicating the outgoing electric field leaving the surface. Figure 3 illustrates the different components of the E vector. For HD-tDCS, the average values of E-normal entering the surface were 1.24E-01 V/m at the 99.9th percentile, 5.89E-02 V/m at the 99th percentile, and 1.16E-02 V/m at the 95th percentile. As for J-normal, the average values entering the surface were 3.40E-02 V/m at the 99.9th percentile, 1.62E-02 V/m at the 99th percentile, and 3.19E-03 V/m at the 95th percentile. Regarding reversed HD-tDCS, the average values of E-normal entering the surface were 1.24E-01 V/m at the 99.9th percentile, 5.89E-02 V/m at the 99th percentile, and 1.16E-02 V/m at the 95th percentile. For J-normal, the average values entering the surface were 8.70E-02 V/m at the 99.9th percentile, 4.01E-02 V/m at the 99th percentile,



and $1.16\text{E-}02$ V/m at the 95^{th} percentile. The focality values for E-normal were $6.38\text{E+}02$ V/m at the 75^{th} percentile and $1.78\text{E+}03$ V/m at the 50^{th} percentile for both HD-tDCS and reversed HD-tDCS. Similarly, the focality values for J-normal were $6.38\text{E+}02$ V/m at the 75^{th} percentile and $1.78\text{E+}03$ V/m at the 50^{th} percentile for both HD-tDCS and reversed HD-tDCS.

The average values of E-normal exiting the surface were $-9.30\text{E-}02$ V/m at the 99.9^{th} percentile, $-4.28\text{E-}02$ V/m at the 99^{th} percentile, and $-1.22\text{E-}02$ V/m at the 95^{th} percentile for HD-tDCS. As for J-normal, the average values entering the surface were $-2.56\text{E-}02$ V/m at the 99.9^{th} percentile, $-1.18\text{E-}02$ V/m at the 99^{th} percentile, and $-3.35\text{E-}03$ V/m at the 95^{th} percentile for HD-tDCS. In contrast, the average values of E-normal entering the surface were $-1.15\text{E-}01$ V/m at the 99.9^{th} percentile, $-5.48\text{E-}02$ V/m at the 99^{th} percentile, and $1.16\text{E-}02$ V/m at the 95^{th} percentile for reversed HD-tDCS. As for J-normal, the average values entering the surface were $8.70\text{E-}02$ V/m at the 99.9^{th} percentile, $4.01\text{E-}02$ V/m at the 99^{th} percentile, and $-1.11\text{E-}02$ V/m at the 95^{th} percentile for reversed HD-tDCS. Regarding focality values, for E-normal, these were $5.84\text{E+}02$ V/m at the 75^{th} percentile and $1.69\text{E+}03$ V/m at the 50^{th} percentile for HD-tDCS, and $6.26\text{E+}02$ V/m at the 75^{th} percentile and $1.76\text{E+}03$ V/m at the 50^{th} percentile for reversed HD-tDCS. As for J-normal, the focality values were $6.38\text{E+}02$ V/m at the 75^{th} percentile and $1.78\text{E+}03$ V/m at the 50^{th}

percentile for both HD-tDCS and reversed HD-tDCS. **Figure 4** illustrates the head models and various components, including E-field, E-normal for HD-tDCS, and reversed HD-tDCS, as well as J-normal for HD-tDCS and reversed HD-tDCS.




Figure 4: The head models and various components, including E-field, E-normal for HD-tDCS, and reversed HD-tDCS, as well as J-normal for HD-tDCS and reversed HD-tDCS. a) Head models using headreco SimNIBS pipeline, b) E-field- Electric field strength across the targeted HD-tDCS placement, c) E-normal at HD-tDCS, d) E-normal at reverse HD-tDCS, e) J-normal at HD-tDCS, f) J-normal at reversed HD-tDCS. Color gradient shows minimum values as blue and maximum values as red.




Figure 5: The electric field strength and current density for all 10 MCI subjects. Peak electric fields for Norm E (a) and Norm J (b) for HD-tDCS and reversed HD-tDCS. Peak electric fields for both Norm E and Norm J changes subject to subject and values are indicated as 99.9, 99 and 95 percentile of total peak values distribution.




Figure 6: Focality for both Norm E and Norm J changes subject to subject and values are indicated as 75 and 50 percentile of total peak values distribution. Focality for norm E (a) and norm J (b).




Figure 7: Peak electric fields for E-normal (a) and J-normal (b) for HD-tDCS.




Figure 8: Peak electric fields for E-normal (a) and J-normal (b) for reversed HD-tDCS.

Table 3: Peak electric field and focality for HD-tDCS and reversed HD-tDCS

Sr. No.	Subject ID	Sex	Age	HD-tDCS and reversed HD-tDCS									
				Peak electric field (V/m)						Focality (mm ³)			
				Norm E (V/m)			Norm J (A/m ²)			Norm E		Norm J	
				99.9	99	95	99.9	99	95	75	50	75	50
1	Sub 1	F	70	1.13E-01	6.42E-02	2.68E-02	3.10E-02	1.76E-02	7.37E-03	1.32E+03	6.71E+03	1.32E+03	6.71E+03
2	Sub 2	M	83	9.44E-02	6.91E-02	2.81E-02	2.60E-02	1.90E-02	7.72E-03	5.36E+03	1.69E+04	5.36E+03	1.69E+04
3	Sub 3	M	79	2.13E-01	1.16E-01	3.52E-02	5.86E-02	3.19E-02	9.69E-03	3.04E+03	7.48E+03	3.04E+03	7.48E+03
4	Sub 4	M	69	2.91E-01	1.67E-01	5.43E-02	8.00E-02	4.60E-02	1.49E-02	2.76E+03	8.29E+03	2.76E+03	8.29E+03
5	Sub 5	F	55	2.46E-01	1.50E-01	4.60E-02	6.78E-02	4.14E-02	1.27E-02	2.61E+03	7.15E+03	2.61E+03	7.15E+03
6	Sub 6	M	75	7.24E-02	4.30E-02	1.99E-02	1.99E-02	1.18E-02	5.49E-03	2.15E+03	9.04E+03	2.15E+03	9.04E+03
7	Sub 7	M	56	1.99E-01	1.20E-01	4.17E-02	5.48E-02	3.29E-02	1.15E-02	3.12E+03	1.04E+04	3.12E+03	1.04E+04
8	Sub 8	M	67	1.36E-01	9.16E-02	3.21E-02	3.74E-02	2.52E-02	8.83E-03	4.36E+03	1.28E+04	4.36E+03	1.28E+04
9	Sub 9	M	77	1.63E-01	8.38E-02	3.66E-02	4.47E-02	2.30E-02	1.01E-02	2.49E+03	7.01E+03	2.49E+03	7.01E+03
10	Sub 10	M	79	9.53E-02	5.98E-02	2.35E-02	2.62E-02	1.65E-02	6.47E-03	2.80E+03	9.71E+03	2.80E+03	9.71E+03
Total			710	1.62E+00	9.65E-01	3.44E-01	4.46E-01	2.65E-01	9.48E-02	3.00E+04	9.55E+04	3.00E+04	9.55E+04
Average			71	1.62E-01	9.65E-02	3.44E-02	4.46E-02	2.65E-02	9.48E-03	3.00E+03	9.55E+03	3.00E+03	9.55E+03

Table 4: Peak electric field and focality for E-normal and J-normal components for HD-tDCS

Sr. No.	Subject ID	Sex	Age	Positive & Negative values	Simulation-1 (HD-tDCS)									
					Peak electric field (V/m)						Focality			
					E- normal (V/m)			J-normal (A/m ²)			E- normal (V/m)		J-normal (A/m ²)	
					99.9	99	95	99.9	99	95	75	50	75	50
1	Sub_1	F	70	P	7.79E-02	4.23E-02	1.17E-02	2.14E-02	1.16E-02	3.23E-03	5.14E+02	1.71E+03	5.14E+02	1.71E+03
				N	-5.43E-02	-2.91E-02	-1.11E-02	-1.49E-02	-8.00E-03	-3.04E-03	4.43E+02	1.70E+03	4.43E+02	1.70E+03
2	Sub_2	M	83	P	7.79E-02	4.36E-02	9.35E-03	2.14E-02	1.20E-02	2.57E-03	8.64E+02	2.21E+03	8.64E+02	2.21E+03
				N	-5.89E-02	-3.15E-02	-9.31E-03	-1.62E-02	-8.67E-03	-2.56E-03	7.28E+02	2.09E+03	7.28E+02	2.09E+03
3	Sub_3	M	79	P	1.63E-01	7.08E-02	1.03E-02	4.49E-02	1.95E-02	2.84E-03	4.53E+02	1.34E+03	4.53E+02	1.34E+03
				N	-1.13E-01	-5.33E-02	-1.24E-02	-3.11E-02	-1.46E-02	-3.41E-03	5.49E+02	1.57E+03	5.49E+02	1.57E+03
4	Sub_4	M	69	P	2.12E-01	9.50E-02	1.79E-02	5.83E-02	2.61E-02	4.93E-03	6.13E+02	1.57E+03	6.13E+02	1.57E+03
				N	-1.75E-01	-6.82E-02	-1.88E-02	-4.82E-02	-1.88E-02	-5.17E-03	4.91E+02	1.17E+03	4.91E+02	1.17E+03
5	Sub_5	F	55	P	1.98E-01	9.24E-02	1.38E-02	5.46E-02	2.54E-02	3.80E-03	6.00E+02	1.40E+03	6.00E+02	1.40E+03
				N	-1.57E-01	-7.15E-02	-1.70E-02	-4.32E-02	-1.97E-02	-4.67E-03	4.92E+02	1.35E+03	4.92E+02	1.35E+03
6	Sub_6	M	75	P	5.65E-02	2.75E-02	7.48E-03	1.55E-02	7.55E-03	2.06E-03	5.45E+02	1.79E+03	5.45E+02	1.79E+03
				N	-4.39E-02	-2.06E-02	-7.11E-03	-1.21E-02	-5.66E-03	-1.96E-03	5.44E+02	1.68E+03	5.44E+02	1.68E+03
7	Sub_7	M	56	P	1.48E-01	6.95E-02	1.19E-02	4.06E-02	1.91E-02	3.28E-03	6.57E+02	1.97E+03	6.57E+02	1.97E+03
				N	-1.04E-01	-4.82E-02	-1.31E-02	-2.86E-02	-1.32E-02	-3.61E-03	7.00E+02	1.93E+03	7.00E+02	1.93E+03
8	Sub_8	M	67	P	1.00E-01	5.24E-02	1.06E-02	2.76E-02	1.44E-02	2.92E-03	8.86E+02	2.32E+03	8.86E+02	2.32E+03
				N	-7.49E-02	-3.64E-02	-1.08E-02	-2.06E-02	-1.00E-02	-2.98E-03	7.26E+02	2.00E+03	7.26E+02	2.00E+03
9	Sub_9	M	77	P	1.21E-01	5.26E-02	1.24E-02	3.33E-02	1.45E-02	3.42E-03	4.65E+02	1.45E+03	4.65E+02	1.45E+03
				N	-8.83E-02	-3.86E-02	-1.24E-02	-2.43E-02	-1.06E-02	-3.42E-03	5.25E+02	1.48E+03	5.25E+02	1.48E+03
10	Sub_10	M	79	P	8.17E-02	4.31E-02	1.04E-02	2.25E-02	1.19E-02	2.87E-03	7.81E+02	2.02E+03	7.81E+02	2.02E+03
				N	-6.06E-02	-3.10E-02	-9.77E-03	-1.67E-02	-8.52E-03	-2.69E-03	6.46E+02	1.89E+03	6.46E+02	1.89E+03

Table 5: Peak electric field and focality for E-normal and J-normal components for reversed HD-tDCS

Sr. No.	Subject ID	Sex	Age	Positive & Negative values	Simulation-2 (Reversed HD-tDCS)									
					Peak electric field (V/m)						Focality			
					E- normal (V/m)			J-normal (A/m ²)			E- normal (V/m)		J-normal (A/m ²)	
					99.9	99	95	99.9	99	95	75	50	75	50
1	Sub_1	F	70	P	5.43E-02	2.91E-02	1.11E-02	1.49E-02	8.00E-03	3.04E-03	4.47E+02	1.70E+03	4.47E+02	1.70E+03
				N	-7.80E-02	-4.23E-02	-1.17E-02	-2.14E-02	-1.16E-02	-3.23E-03	5.14E+02	1.71E+03	5.14E+02	1.71E+03
2	Sub_2	M	83	P	5.89E-02	3.15E-02	9.31E-03	1.62E-02	8.66E-03	2.56E-03	7.30E+02	2.09E+03	7.30E+02	2.09E+03
				N	-7.79E-02	-4.36E-02	-9.35E-03	-2.14E-02	-1.20E-02	-2.57E-03	8.63E+02	2.21E+03	8.63E+02	2.21E+03
3	Sub_3	M	79	P	1.13E-01	5.33E-02	1.24E-02	3.11E-02	1.46E-02	3.41E-03	5.49E+02	1.57E+03	5.49E+02	1.57E+03
				N	-1.63E-01	-7.09E-02	-1.03E-02	-4.50E-02	-1.95E-02	-2.84E-03	4.53E+02	1.34E+03	4.53E+02	1.34E+03
4	Sub_4	M	69	P	1.75E-01	6.82E-02	1.88E-02	4.81E-02	1.87E-02	5.17E-03	4.91E+02	1.17E+03	4.91E+02	1.17E+03
				N	-2.12E-01	-9.50E-02	-1.79E-02	-5.84E-02	-2.61E-02	-4.93E-03	6.12E+02	1.57E+03	6.12E+02	1.57E+03
5	Sub_5	F	55	P	1.57E-01	7.14E-02	1.70E-02	4.31E-02	1.96E-02	4.67E-03	4.92E+02	1.35E+03	4.92E+02	1.35E+03
				N	-1.99E-01	-9.26E-02	-1.38E-02	-5.46E-02	-2.55E-02	-3.80E-03	5.98E+02	1.39E+03	5.98E+02	1.39E+03
6	Sub_6	M	75	P	4.39E-02	2.06E-02	7.11E-03	1.21E-02	5.65E-03	1.96E-03	5.44E+02	1.68E+03	5.44E+02	1.68E+03
				N	-5.65E-02	-2.75E-02	-7.49E-03	-1.55E-02	-7.55E-03	-2.06E-03	5.45E+02	1.79E+03	5.45E+02	1.79E+03
7	Sub_7	M	56	P	4.39E-02	2.06E-02	7.11E-03	1.21E-02	5.65E-03	1.96E-03	5.44E+02	1.68E+03	5.44E+02	1.68E+03
				N	-5.65E-02	-2.75E-02	-7.49E-03	-1.55E-02	-7.55E-03	-2.06E-03	5.45E+02	1.79E+03	5.45E+02	1.79E+03
8	Sub_8	M	67	P	7.49E-02	3.64E-02	1.08E-02	2.06E-02	1.00E-02	2.98E-03	7.28E+02	2.00E+03	7.28E+02	2.00E+03
				N	-1.00E-01	-5.24E-02	-1.06E-02	-2.76E-02	-1.44E-02	-2.92E-03	8.84E+02	2.31E+03	8.84E+02	2.31E+03
9	Sub_9	M	77	P	8.83E-02	3.86E-02	1.24E-02	2.43E-02	1.06E-02	3.42E-03	5.25E+02	1.48E+03	5.25E+02	1.48E+03
				N	-1.22E-01	-5.26E-02	-1.24E-02	-3.34E-02	-1.45E-02	-3.42E-03	4.63E+02	1.44E+03	4.63E+02	1.44E+03
10	Sub_10	M	79	P	6.06E-02	3.10E-02	9.77E-03	1.67E-02	8.52E-03	2.69E-03	6.46E+02	1.89E+03	6.46E+02	1.89E+03
				N	-8.18E-02	-4.31E-02	-1.04E-02	-2.25E-02	-1.19E-02	-2.87E-03	7.80E+02	2.01E+03	7.80E+02	2.01E+03

DISCUSSION

The data obtained from simulations showed variation among all subjects with the same stimulation protocols due to individual differences in age and head geometry. There appears to be no correlation between age and head geometry. The calculated distance for the most aged subject between the nasion andinion was 227.33 mm, while the distance between the right pre-auricular point (RPA) and left pre-auricular point (LPA) was 159.97 mm. The average distance between the inion and nasion, based on fiducial points, was 200.9452 mm, and between RPA and LPA was 152.2565 mm. Additionally, the distance between the central electrode and the surrounding four electrodes varied, but there were no correlations between head geometry and the corresponding electric field strength with their E-normal and J-normal components. **Figure 4** and **Figure 5** display the electric field strength and current density for all 10 MCI subjects. Among all subjects, significantly high electric field values (norm-E) were observed using HD-tDCS and reversed HD-tDCS, with values of $1.62\text{E-}01$ V/m at the 99.9th percentile, $9.65\text{E-}02$ V/m at the 99th percentile, and $3.44\text{E-}02$ V/m at the 95th percentile. Similarly, the current density values followed a decreasing order at the 99.9th (4.46E-02 A/m²), 99 (2.65E-02 A/m²), and 95th (9.48E-03 A/m²) percentiles. However, no significant relationships or dependencies were observed between age, head geometry, and the distance between electrodes in MCI subjects, as the progression of the disease was unknown. Contrasting the values of norm E and norm J, the peak values for focality showed an increasing order at the 75th and 50th percentiles, with values of $3.00\text{E+}03$ V/m and $9.55\text{E+}03$ V/m for norm E and $3.00\text{E+}03$ V/m and $9.55\text{E+}03$ V/m for norm J, as depicted in **Figure 6a** and **Figure 5 b**.

The average normal values of the E vector were consistently lower than the total E-field obtained, with values of $1.24\text{E+}00$ V/m, $5.89\text{E-}01$ V/m, and $1.16\text{E-}01$ V/m for the 99.5th, 99th, and 95th percentiles, respectively, for E-normal (**Figure 7a** and **Figure 6 b**). The same applied to the current density. In HD-tDCS, there were more positive values, indicating a greater inflow of voltage and current compared to reversed HD-tDCS. Conversely, in reversed HD-tDCS, there were more negative values, suggesting a higher outflow of current and voltage compared to HD-tDCS, (**Figure 8 a** and **Figure 7 b**).

CONCLUSION

The electric field is affected by multiple factors, including electrode size, shape, tissue conductivity, and

current distribution pattern. A smaller diameter of the electrode generates a more focused electric field distribution. The distribution of the electric field in the brain is intricate and relies on various factors, including the electrode arrangement and the head's conductivity profile. It is essential to understand that the specific current density and stimulation parameters utilized in tDCS can differ based on the particular study, the targeted brain region, and individual characteristics. HD-tDCS and reversed HD-tDCS exhibit distinct values for electric field strength and focality. In HD-tDCS, we observed more positive fields and inflow current compared to reversed HD-tDCS, making it suitable for excitatory effects. Conversely, reversed HD-tDCS is more appropriate for inhibitory effects due to the prevalence of more negative fields and outflow current. In this study, we focused specifically on individuals with MCI. The outcomes may vary among subjects. Additional subjects can be included to further investigate the uniformity and variance of the electric field and focality for MCI or other neurological disorders.

ABBREVIATIONS

3D: Three dimensional, **AD**: Alzheimer's disease, **CAT**: Computational anatomy toolbox, **COMET**: Computation of Electric field due to Transcranial current Stimulation, **EEG**: Electroencephalograph, **GR**: Gradient (Pulse Sequence), **HD-tDCS**: High-Definition tDCS, **Iz**: Inion, **LPA**: Left pre-auricle, **mA**: Milli ampere, **MCI**: Mild Cognitive Impairment, **MNI**: Montreal Neurologic Institute, **MoCA**: Montreal Cognitive Assessment, **MRI**: Magnetic Resonance Imaging, **Nz**: Nasion, **PET**: Positron Emission Tomography, **ROAST**: A fully automated, Realistic, volumetric approach to simulate transcranial electric stimulation, **RPA**: Right pre-auricle, **SimNIBS**: Simulation of Non-invasive brain stimulation, **SPM**: Statistical parametric mapping, **tDCS**: Transcranial direct current stimulation, **TE**: Echo Time, **TI**: Inversion Time, **TR**: Repetition Time, **ts-DCS**: Transcutaneous spinal direct current stimulation, **V/m**: Volts per meter

ACKNOWLEDGMENTS

We would like to express our sincere gratitude to the reviewers for their valuable feedback and insightful comments on our research paper. Their constructive criticism and thoughtful suggestions have greatly contributed to the improvement and refinement of this manuscript. We also thank the editorial team for their meticulous handling of our submission and for facilitating the peer review process.

AUTHOR'S CONTRIBUTIONS

Utkarsh Pancholi conducted the experiments, collected and analyzed the data, and drafted the initial manuscript. Both authors contributed to the interpretation of the results, provided critical revisions, and played a key role in finalizing the manuscript. As a corresponding author Vijay Dave has coordinated the collaboration between co-authors, ensuring everyone was informed of the manuscript's status, revisions needed, and deadlines, thereby fostering a cohesive team effort. All authors read and approved the final manuscript.

FUNDING

None.

AVAILABILITY OF DATA AND MATERIALS

Data and materials used and/or analyzed during the current study are available from the corresponding author on reasonable request.

ETHICS APPROVAL AND CONSENT TO PARTICIPATE

Not applicable.

CONSENT FOR PUBLICATION

Not applicable.

COMPETING INTERESTS

The authors declare that they have no competing interests.

REFERENCES

- Ageing and health [Internet]. [cited 2023 Jul 4]. Available from: <https://www.who.int/news-room/fact-sheets/detail/ageing-and-health>.
- Parsaik AK, Singh B, Roberts RO, Pankratz S, Edwards KK, Geda YE. Hypothyroidism and risk of mild cognitive impairment in elderly persons: a population-based study. *JAMA Neurology*. 2014;71(2):201–7. PMID: 24378475. Available from: <https://doi.org/10.1001/jamaneurol.2013.5402>.
- Wang Y, Pan F, Xie F, He R, Guo Q. Correlation Between Urine Formaldehyde and Cognitive Abilities in the Clinical Spectrum of Alzheimer's Disease. *Frontiers in Aging Neuroscience*. 2022;14. PMID: 35221998. Available from: <https://doi.org/10.3389/fnagi.2022.820385>.
- Papaliagkas V, Kalinderi K, Varelziz P, Moraitou D, Papamitsou T, Chatzidimitriou M. CSF Biomarkers in the Early Diagnosis of Mild Cognitive Impairment and Alzheimer's Disease. *International Journal of Molecular Sciences*. 2023;24(10):8976. Available from: <https://doi.org/10.3390/ijms24108976>.
- Chandra A, Dervenoulas G, Politis M, Initiative ADN. Magnetic resonance imaging in Alzheimer's disease and mild cognitive impairment. *Journal of Neurology*. 2019;266(6):1293–302. PMID: 30120563. Available from: <https://doi.org/10.1007/s00415-018-9016-3>.
- Teng L, Li Y, Zhao Y, Hu T, Zhang Z, Yao Z, et al. Predicting MCI progression with FDG-PET and cognitive scores: a longitudinal study. *BMC Neurology*. 2020;20(1):148. PMID: 32316912. Available from: <https://doi.org/10.1186/s12883-020-01728-x>.
- Narayanan L, Murray AD. What can imaging tell us about cognitive impairment and dementia? *World Journal of Radiology*. 2016;8(3):240–54. PMID: 27029053. Available from: <https://doi.org/10.4329/wjr.v8.i3.240>.
- Hobson J. The Montreal Cognitive Assessment (MoCA). *Occupational Medicine (Oxford, England)*. 2015;65(9):764–5. PMID: 26644445. Available from: <https://doi.org/10.1093/occmed/kqv078>.
- Petersen RC, Lopez O, Armstrong MJ, Getchius TS, Ganguli M, Gloss D. Practice guideline update summary: Mild cognitive impairment: Report of the Guideline Development, Dissemination, and Implementation Subcommittee of the American Academy of Neurology. *Neurology*. 2018;90(3):126–35. PMID: 29282327. Available from: <https://doi.org/10.1212/WNL.0000000000004826>.
- Chen YX, Liang N, Li XL, Yang SH, Wang YP, Shi NN. Diagnosis and Treatment for Mild Cognitive Impairment: A Systematic Review of Clinical Practice Guidelines and Consensus Statements. *Frontiers in Neurology*. 2021;12. PMID: 34712197. Available from: <https://doi.org/10.3389/fneur.2021.719849>.
- Ma L. Depression, Anxiety, and Apathy in Mild Cognitive Impairment: current Perspectives. *Frontiers in Aging Neuroscience*. 2020;12:9. PMID: 32082139. Available from: <https://doi.org/10.3389/fnagi.2020.00009>.
- Tuokko H, Frerichs RJ. Cognitive impairment with no dementia (CIND): longitudinal studies, the findings, and the issues. *The Clinical Neuropsychologist*. 2000;14(4):504–25. PMID: 11262720. Available from: <https://doi.org/10.1076/clin.14.4.504.7200>.
- Costain R, Redfearn JW, Lippold OC. A controlled trial of the therapeutic effects of polarization of the brain in depressive illness. *The British Journal of Psychiatry*. 1964;110(469):786–99. PMID: 14211695. Available from: <https://doi.org/10.1192/bjp.110.469.786>.
- Carney MW. Negative polarisation of the brain in the treatment of manic states. *Irish Journal of Medical Science*. 1969;8(3):133–5. PMID: 5799726. Available from: <https://doi.org/10.1007/BF02958921>.
- Pancholi U, Dave V. Analysis of electric field strengths and focality for healthy and neurologically impaired subjects upon multiple tDCS stimulation protocols. *Research on Biomedical Engineering*. 2023;39(4):991–1006. Available from: <https://doi.org/10.1007/s42600-023-00322-7>.
- Purpura DP, McMurtry JG. Intracellular activities and evoked potential changes during polarization of motor cortex. *Journal of Neurophysiology*. 1965;28(1):166–85. PMID: 14244793. Available from: <https://doi.org/10.1152/jn.1965.28.1.166>.
- Pancholi U, Dave V. Quantification of electric field strength of tDCS in Alzheimer's and mild cognitive impairment patients. *Int J Adv Technol Eng Explor*. 2021;8(82):1168–82. Available from: <https://doi.org/10.19101/IJATEE.2021.874444>.
- Pancholi U, Dave V. Variability of E-field in Dorsolateral Prefrontal Cortex Upon a Change in Electrode Parameters in tDCS. In: *International Conference on Bio Signals, Images, and Instrumentation (ICBSII)*; 2023. p. 1–8. Available from: <https://doi.org/10.1109/ICBSII58188.2023.10180905>.
- Nitsche MA, Cohen LG, Wassermann EM, Priori A, Lang N, Antal A. Transcranial direct current stimulation: state of the art 2008. *Brain Stimulation*. 2008;1(3):206–23. PMID: 20633386. Available from: <https://doi.org/10.1016/j.brs.2008.06.004>.
- Datta A, Bansal V, Diaz J, Patel J, Reato D, Bikson M. Gyri-precise head model of transcranial direct current stimulation: improved spatial focality using a ring electrode versus conventional rectangular pad. *Brain Stimulation*. 2009;2(4):201–7. PMID: 20648973. Available from: <https://doi.org/10.1016/j.brs.2009.03.005>.
- Sharma G, Chowdhury SR. *IEEE Conf Comput Intell Bioinforma Comput Biol CIBCB* 2019; 2019.

22. Morya E, Monte-Silva K, Bikson M, Esmaeilpour Z, Biazoli CE, Fonseca A, et al. Beyond the target area: an integrative view of tDCS-induced motor cortex modulation in patients and athletes. *Journal of neuroengineering and rehabilitation*. 2019;16(1):1–29. Available from: <https://doi.org/10.1186/s12984-019-0581-1>.
23. Pancholi UV, Dave V. Review of computational approaches to model transcranial direct current stimulations tDCS and its effectiveness. *J Integr Sci Technol*. 2022;10(1):1–10.
24. Lefaucheur JP, André-Obadia N, Antal A, Ayache SS, Baeken C, Benninger DH. Evidence-based guidelines on the therapeutic use of repetitive transcranial magnetic stimulation (rTMS). *Clinical Neurophysiology*. 2014;125(11):2150–206. PMID: 25034472. Available from: <https://doi.org/10.1016/j.clinph.2014.05.021>.
25. Fu Q, Cheng QG, Kong XC, Liu DX, Guo YH, Grinstead J. Comparison of contrast-enhanced T1-weighted imaging using DANTE-SPACE, PETRA, and MPRAGE: a clinical evaluation of brain tumors at 3 Tesla. *Quantitative Imaging in Medicine and Surgery*. 2022;12(1):592–607. PMID: 34993104. Available from: <https://doi.org/10.21037/qims-21-107>.
26. Habes M, Sotiras A, Erus G, Toledo JB, Janowitz D, Wolk DA. White matter lesions: spatial heterogeneity, links to risk factors, cognition, genetics, and atrophy. *Neurology*. 2018;91(10):e964–75. PMID: 30076276. Available from: <https://doi.org/10.1212/WNL.0000000000006116>.
27. ADNI | Alzheimer's Disease Neuroimaging Initiative [Internet]. [cited 2022 Nov 14]. Available from: <https://adni.loni.usc.edu/>.
28. Saturnino GB, Puonti O, Nielsen JD, Antonenko D, Madsen KH, Thielscher A. A Comprehensive Pipeline for Individualized Electric Field Modelling for Transcranial Brain Stimulation. *bioRxiv*. 2018;p. 1–13. Available from: <https://doi.org/10.1101/500314>.
29. Saturnino GB, Madsen KH, Thielscher A. Efficient Electric Field Simulations for Transcranial Brain Stimulation. *bioRxiv*. 2019;p. 541409. Available from: <https://doi.org/10.1101/541409>.
30. Mikkonen M, Laakso I, Tanaka S, Hirata A. Cost of focality in TDCS: interindividual variability in electric fields. *Brain Stimulation*. 2020;13(1):117–24. PMID: 31606449. Available from: <https://doi.org/10.1016/j.brs.2019.09.017>.